

Artificial Neural Networks for Recognition of Electrocardiographic Lead Reversal

Bo Hedén, MD, Mattias Ohlsson, MSc, Lars Edenbrandt, MD, PhD, Ralf Rittner, MSc, Olle Pahlm, MD, PhD, and Carsten Peterson, PhD

Misplacement of electrodes during the recording of an electrocardiogram (ECG) can cause an incorrect interpretation, misdiagnosis, and subsequent lack of proper treatment. The purpose of this study was twofold: (1) to develop artificial neural networks that yield peak sensitivity for the recognition of right/left arm lead reversal at a very high specificity; and (2) to compare the performances of the networks with those of 2 widely used rule-based interpretation programs. The study was based on 11,009 ECGs recorded in patients at an emergency department using computerized electrocardiographs. Each of the ECGs was used to computationally generate an ECG with right/left arm lead reversal. Neural networks were trained to detect ECGs with right/left arm lead reversal. Different networks and rule-based criteria were used depending on the presence or absence of P waves. The networks and the cri-

teria all showed a very high specificity (99.87% to 100%). The neural networks performed better than the rule-based criteria, both when P waves were present (sensitivity 99.1%) or absent (sensitivity 94.5%). The corresponding sensitivities for the best criteria were 93.9% and 39.3%, respectively. An estimated 300 million ECGs are recorded annually in the world. The majority of these recordings are performed using computerized electrocardiographs, which include algorithms for detection of right/left arm lead reversals. In this study, neural networks performed better than conventional algorithms and the differences in sensitivity could result in 100,000 to 400,000 right/left arm lead reversals being detected by networks but not by conventional interpretation programs.

(Am J Cardiol 1995;75:929-933)

Both clinicians and interpretation programs require correct data for a correct electrocardiographic interpretation. Misplacement of electrodes during the electrocardiographic recording is one situation that results in incorrect data. Treatment errors due to lead reversal have been reported; misdiagnosis and subsequent lack of proper treatment as well as inappropriate treatment due to false-positive diagnoses do occur.¹ To prevent a false interpretation, the misplacement should be recognized and corrected by the technician who records the electrocardiogram (ECG). To recognize the lead reversal from the appearance of the ECG may be difficult. Even trained ECG readers often fail to recognize a lead reversal.² Therefore, interpretation programs contain algorithms for detecting the most common type of lead reversal, the right/left arm lead reversal. Two widely used algorithms for the detection of right/left arm lead reversal are known to have a high specificity but rather low sensitivity. Although it is easy for the experienced ECG reader to detect an ECG with a lead reversal, improvement in the algorithms, which are rule-based, is a difficult task, even for the expert. Artificial neural networks (ANNs) have proved to be very powerful in pattern recognition tasks because they can handle almost any nonlinear dependence. ANNs mimic, in a very crude way, mammal information processing. Input data are

processed through layers of artificial neurons connected by weights (Figure 1). ANNs have been used for classification of ECGs.³⁻⁵ In 1 study, an ANN diagnosed myocardial infarction from ECGs better than a conventional interpretation program,⁶ and in another study an ANN performed on par with an experienced ECG reader.⁷ The purpose of this study was twofold: (1) to develop ANNs that yield high sensitivity for the recognition of right/left arm lead reversal at a very high specificity; and (2) to compare the performances of the ANNs with those of 2 widely used interpretation programs.

METHODS

Study population: The study was based on 11,432 ECGs recorded in patients who presented at the emergency ward at the University hospital in Lund during 1992 and 1993. The 12-lead ECGs were recorded using computerized electrocardiographs (Siemens-Elema AB, Solna, Sweden). All recordings were obtained digitally and averaged heart cycles were calculated. The P, QRS, and ST-T measurements used in the criteria and as inputs to the ANNs were obtained from the measurement program of the computerized electrocardiographic recorders.

As stated, the purpose of the study was to develop ANNs that detect right/left arm lead reversal with a very high specificity. Because ANNs learn from example, it was important that no ECG with right/left arm lead reversal was used as an example of a correctly recorded ECG, and vice versa. Therefore, great care was taken to exclude ECGs with lead reversal as well as ECGs that were technically deficient. In addition, pacemaker ECGs were excluded because QRS morphology or the ST-T segment are not assessed by the programs in these ECGs. The exclusion process included both visual inspection of the ECGs and computerized methods. Two experienced ECG readers examined the ECGs independently. The com-

From the Departments of Clinical Physiology and Theoretical Physics, Lund University, Lund, Sweden. This study was supported in part by grants from the Swedish Medical Research Council (B94-14VV-09893-03A), Swedish National Board for Industrial and Technical Development, the Faculty of Medicine at Lund University, the Göran Gustafsson Foundation for Research in National Science and Medicine, and the Swedish Natural Science Research Council, Stockholm, Sweden. Manuscript received November 10, 1994; revised manuscript received and accepted February 7, 1995.

Address for reprints: Bo Hedén, MD, Department of Clinical Physiology, University Hospital, S-221 85 Lund, Sweden.

TABLE I Reasons for Exclusion of Electrocardiograms


Reason for Exclusion	Number of ECGs
Right/left arm lead reversal	47
Right arm/foot lead reversal	32
Left arm/foot lead reversal	8
Precordial lead misplacement	25
Pacemaker ECG	197
Technically deficient ECG	114
Total	423

ECG = electrocardiogram.

TABLE II Measurements Used in the GRI and Marquette Criteria and in the Artificial Neural Networks

Measurement	P Waves Present			P Waves Absent	
	GRI	Marquette	Best ANN	GRI	Marquette
P axis	x	x	x		
P sum (I, V ₆)			x		
P+, P- (V ₆)	x				
QRS axis	x	x	x	x	x
QRS area (I)	x		x	x	
QRS area (V ₆)	x		x		
Q, R ampl (I)	x			x	x
S, R', ST, T+, T- ampl. (I) R, S ampl. (V ₆)	x			x	
QRS PP (V ₆)	x				
QRS ampl. (35, 40, 45), T sum (I, V ₆)			x		
ST, T+, T- ampl. (V ₆)	x			x	

ANN = artificial neural network; PP = peak-to-peak amplitude; QRS ampl. (35, 40, 45) = amplitude 35, 40, and 45 ms after QRS onset; sum = maximal positive amplitude - |maximal negative amplitude|; + = maximal positive amplitude; - = maximal negative amplitude.

**FIGURE 1.** Schematic diagram of a neural network with 15, 5, and 1 neurons in the input, hidden, and output layers, respectively. The neurons are connected to each other by the lines.


puterized methods included histogramming all the variables and a more elaborate method of using an ANN density map.

Technically deficient recordings and pacemaker ECGs were excluded. Based on visual and computerized analysis, ECGs with suspected electrode misplacement or lead reversal were scrutinized by the 2 experts together. Electrode misplacement or lead reversal was verified in most cases after a comparison between the suspected ECG

and a previous or later recording from the same patient found in the clinical database. A total of 423 ECGs were excluded (Table I), which left 11,009 ECGs for the study.

Each of these 11,009 ECGs was used to generate an ECG with right/left arm lead reversal. This was performed computationally by means of inverting lead I, interchanging leads II and III, and interchanging leads aVL and aVR (note that aVF is not affected by the right/left arm lead reversal) (Figure 2). This yielded exactly the same ECG that would have resulted if the right and left arm electrodes had been switched in the recording situation. Thus, the final material consisted of 22,018 ECGs divided into 2 groups; 11,009 ECGs recorded with correct electrode placement and 11,009 with right/left arm lead reversal.

The interpretation programs use different criteria depending on the presence/absence of P waves. Similarly, different ANNs were used if P waves were present or absent. Therefore, only 9,296 pairs of ECGs with P waves could be used to test criteria and ANNs

**FIGURE 2.** A correctly recorded electrocardiogram (ECG) (left) with the corresponding right/left arm lead reversal (right). The result of the computer-generated lead reversal is the same as an electrode misplacement, namely an inverted lead I, a switch of leads II and III, and a switch of leads aVR and aVL.

that use P-wave information. However, all ECGs were used in the analysis of criteria and ANNs not requiring P-wave data as inputs. This approach is justified since the QRS patterns of right/left arm lead reversal are not dependent on the P waves.

Conventional criteria: Performances of the neural networks were compared with those of 2 conventional interpretation programs, namely the GRI program developed at the Glasgow Royal Infirmary⁸ and the Marquette program.⁹ Both programs use rule-based criteria. For example, if P waves are detected, the following rule is used in the Marquette program: If the QRS axis is between 90° and 270° and the P axis is between 90° and 210°, then say "suspect arm lead reversal."

All ECGs were processed by the Glasgow measurement program and these measurements were used as inputs to both interpretation programs. The inputs to the 2 programs are presented in Table II.

Neural network: A multilayer perceptron ANN architecture¹⁰ and Langevin updating procedure¹¹ were used. A more general description of ANN can be found elsewhere.¹² The ANNs consisted of 1 input layer, 1 hidden layer, and 1 output layer (Figure 1). The output unit encodes whether the ECG is correctly recorded (output value = 1) or if a right/left arm lead reversal is detected (output value = 0). The hidden layer contained 4 to 6 neurons. The number of neurons in the input layer equals the number of input variables. These are presented in Table II for the 5 different networks. The P and QRS axes were presented to the networks as $\sin(\text{axis} \cdot \pi/180)$ and $\cos(\text{axis} \cdot \pi/180)$. In cases in which the P or QRS axis was undetermined, both the sin and cos values were set to 0.

Similar to conventional criteria, different ANNs were used if P waves were or were not present. Two networks used the same inputs as the GRI program, whereas 2 other networks used the same inputs as the Marquette program. Furthermore, selections of the available inputs were used to find ANNs that yielded high performance.

The data set was divided into 2 parts: a training set and a test set. The training set was used to adjust the connection weights, whereas the test set was used to assess the performance. To get as reliable a performance as possible, a K-fold cross validation was used in which the data set was randomly divided into K equal parts. For each of the K different test sets, training was performed on the remaining (K-1)/K parts of the data. We used 3-fold cross validation to decide when to terminate learning in order to avoid "overtraining," and 7-fold cross

TABLE III Results for the GRI and Marquette Criteria and for the Artificial Neural Networks

Criteria/ANN	Specificity (%)	Sensitivity (%)
P waves present		
GRI criteria	100.00	84.02
ANN using GRI measurements	99.94 ± 0.026	98.7 ± 0.12
Marquette criteria	99.87	93.91
ANN using Marquette measurements	99.92 ± 0.022	94.8 ± 0.26
Best ANN	99.95 ± 0.017	99.11 ± 0.080
P waves absent		
GRI criteria	100.00	39.34
ANN using GRI measurements	99.92 ± 0.023	94.5 ± 0.52
Marquette criteria	99.91	30.93
ANN using Marquette measurements	99.90 ± 0.028	63 ± 2.6

Results are expressed as mean ± SD.
Abbreviations as in Table II.




FIGURE 3. A correctly recorded electrocardiogram from a 50-year-old patient with severe lung and heart disease. The electrocardiogram was falsely classified as right/left arm lead reversal by the artificial neural network and the Marquette program.

validation to train the networks and assess their performances. The error estimates in Table III result from 25 independent runs for each type of network.

During the training process, the connection weights between the neurons were adjusted using the back-propagation algorithm. The learning rate (η) had a start value of 0.5. During the training, η was decreased geometrically every epoch using the following equation: $\eta = k \eta$ with $k = 0.998$.

The momentum α was set to 0.7. Updating occurred after each 20 patterns. The Langevin noise was chosen to decrease geometrically from 0.005 and with $k = 0.993$ during the training process. The network weights were initiated with random numbers between -0.025 and 0.025. Parameters that were separate for the different ANNs are presented in Table IV. All calculations were done using the JETNET 3.0 package.¹³

To achieve as high a specificity as possible, at the cost of a lower sensitivity, the following 2 methods were used during the training procedure: (1) The correctly recorded ECGs were presented to the ANN typically 50

times more often than the incorrect ECGs. (2) An asymmetric error function was used (i.e., the network was penalized more for a false-positive case than for a false-negative case).

RESULTS

The results for the conventional criteria and the ANNs are presented in Table III. All criteria and all ANNs show very high specificity. One of the few correct ECGs classified as right/left arm lead reversal by the ANNs and the Marquette criteria is shown in Figure 3. This ECG was also classified as possible lead reversal by both experts independently. After a comparison with previous ECGs from the patient in a clinical database, it was shown that the ECG was correct and the atypical complex in lead I was due to severe lung and heart disease.

In the presence of P waves, conventional criteria show high sensitivity. However, with use of the ANNs, the sensitivity was even higher. Figure 4A shows an example of right/left arm lead reversal that was missed by the GRI criteria. The networks correctly detected the ECG as a lead reversal, which is obvious for the experienced ECG reader. The conventional criteria had a much lower sensitivity in the absence of P-wave data: 30.9% and 39.3% for the Marquette and GRI criteria, respectively. The corresponding sensitivity for the ANN, which used the same variables as the GRI criteria, was 94.5%. One example of lead reversal missed by both criteria and detected by the ANNs is shown in Figure 4B.

To check to what extent our database was sufficient in size, the ANNs were trained and tested using half of the material. No significant difference was observed.

DISCUSSION

One objective of the present study was to develop ANNs with high sensitivity for the recognition of right/left arm lead reversal at very high specificity. The importance of high specificity could be illustrated by the following example: The specificity of the best networks were as high as 99.95% (i.e., 1 false-positive ECG of 2,000 ECGs). A right/left arm lead reversal occurs only in 1 out of 100 to 400 ECGs depending on the experience of the recording technician. In considering a set of 2,000 ECGs, 5 to 20 of these will have a right/left arm lead reversal, and 1 correct ECG will falsely be reported as incorrect using this neural network. With a sensitivity of approximately 95%, most of the incorrect ECGs will also be reported. Consequently, most of the ECGs that the neural network classifies as right/left arm lead reversal will actually be a case with lead reversal (i.e., the positive predictive value will be high).

Ideally, the network would never report lead reversal for a correct ECG. The network learns from example and it is therefore very important that the training set does not include such examples (i.e., a correct ECG labeled as a lead reversal). This would happen if an ECG with right/left arm lead reversal was included in the data set of 11,009 ECGs labeled as correct. Therefore, great care must be taken in purification of the database. In this study, ANNs were used as a complement to the visual examination of the ECG experts in this process.

The performance of an ANN is dependent on the size and composition of the database used for training. Therefore the ECGs of the training set should be representative for routine ECGs processed by computerized ECG recorders. To accomplish this, a database of >11,000 ECGs




FIGURE 4. Two electrocardiograms (ECGs) with right/left arm lead reversal, 1 with P waves present (A) and 1 without P waves (B). Both ECGs were detected by artificial neural networks, whereas the GRI criteria missed them and the Marquette criteria missed the ECG without P waves.

was selected. Training and testing ANNs using only half the material did not impair the performance, which shows that the size of the material was sufficient. The ECGs used to train the ANNs were recorded at an emergency ward (i.e., in a clinical setting where the ANNs could be of greatest help).

A second purpose of the study was to compare the performances of the ANNs with those of 2 widely used interpretation programs. Both programs and ANNs showed very high specificities, but sensitivities were much higher for the ANNs. Considering that over 100 million ECGs are recorded annually in the U.S.¹⁴ and probably another 200 million in the rest of the world, this difference in sensitivity could result in 100,000 to 400,000 right/left arm lead reversals being detected by ANNs but not by conventional interpretation programs. It has been shown that even cardiologists fail to recognize right arm/leg lead reversal.² The right/left arm lead reversal is probably also overlooked by many ECG readers. Therefore, the detection of lead reversal in interpretation programs is a very important type of quality control.

1. Guijarro-Morales A, Gil-Extremera B, Maldonado-Martin A. ECG diagnostic errors due to improper connection of the right arm and leg cable. *Int J Cardiol* 1991;30:233-235.
2. Haisty WK Jr, Pahlm O, Edenbrandt L, Newman K. Recognition of electrocardiographic electrode misplacements involving the ground (right leg) electrode. *Am J Cardiol* 1993;71:1490-1495.
3. Edenbrandt L, Devine B, Macfarlane PW. Neural networks for classification of electrocardiographic ST-T segments. *J Electrocardiol* 1992;25:167-173.
4. Edenbrandt L, Devine B, Macfarlane PW. Classification of electrocardiographic ST-T segments—human expert versus artificial neural network. *Eur Heart J* 1993;14:464-468.

TABLE IV Network Structure for Different Input Data Set

Neural network	Number of Neurons			SR	AE	Err _{min}	<Epochs>
	Input	Hidden	Output				
P waves present							
ANN using GRI measurements	21	5	1	40:1	1	0.020	148
ANN using Marquette measurements	4	4	1	70:1	1	0.030	325
Best ANN	16	5	1	20:1	10	0.022	125
P waves absent							
ANN using GRI measurements	15	5	1	70:1	1	0.030	170
ANN using Marquette measurements	4	6	1	50:1	5	0.075	500

AE = the β parameter in the asymmetric error function (AE = 1 is the normal squared error function); <Epochs> = the average number of epochs, where 1 epoch is completed when the same number of patterns have been presented to the network as the number of patterns in the training set; Err_{min} = error level in the training set at which the training is stopped; SR = sampling ratio between correctly recorded electrocardiograms and electrocardiograms with right/left arm lead reversal; other abbreviations as in Table II.

5. Bortolan G, Degani R, Willems JL. Neural networks for ECG classification. In: *Computers in Cardiology 1990*. Los Alamitos, CA: IEEE Computer Society Press, 1990:269-272.
6. Hedén B, Edenbrandt L, Haisty WK Jr, Pahlm O. Artificial neural networks for the electrocardiographic diagnosis of healed myocardial infarction. *Am J Cardiol* 1994;74:5-8.
7. Reddy MRS, Edenbrandt L, Svensson J, Haisty WK, Pahlm O. Neural network versus electrocardiographer and conventional computer criteria in diagnosing anterior infarct from the ECG. In: *Computers in Cardiology 1992*. Los Alamitos, CA: IEEE Computer Society Press, 1992:667-670.
8. Macfarlane PW, Lawrie TDV. *Comprehensive Electrocardiology*. vol 3. Oxford: Pergamon Press, 1989:1529-1530.
9. *Physicians Guide to Marquette Electronics Resting ECG Analysis*. Milwaukee, WI: Marquette Electronics, 1988:53.
10. Rumelhart DE, McClelland JL, eds. *Parallel Distributed Processing*. Vol. 1 & 2. Cambridge, MA: MIT Press, 1986.
11. Rögnvaldsson T. On Langevin updating in multilayer perceptrons. *Neural Computation* 1994;6:916-926.
12. Hertz J, Krogh A, Palmer RG. *Introduction to the Theory of Neural Computation*. Redwood City, CA: Addison-Wesley, 1991.
13. Peterson C, Rögnvaldsson T, Lönnblad L. JETNET 3.0—a versatile artificial neural network package. *Comput Phys Commun* 1994;81:185-220.
14. Drazen E, Mann N, Borun R, Laks M, Berson A. Survey of computer-assisted electrocardiography in the United States. *J Electrocardiol* 1988;21:S98-S104.